

EU Type Examination Certificate

This is to certify that:

Mercator Medical S. A.
Ul. H. Modrzejewskiej 30
Cracow
31-327
Poland

Holds Certificate Number:

CE 698940

In respect of:

Gloves for personal protection. Model: Ambulance high risk - Powder Free Latex Gloves.

on the basis that BSI carried out the relevant Type Examination procedures under the requirements with the Regulation (EU) 2016/425 of the European Parliament and Council relating to Personal Protective Equipment Regulation (PPE) Annex V (Module B) and meets the relevant health and safety requirements specified in Annex II

For and on behalf of BSI, a Notified Body for the above Regulation (Notified Body Number 2797):

Denelise L'Ecluse, Managing Director Assurance - Continental Europe

First Issued: 2018-11-06

Latest Issue: 2023-10-09

Effective Date: 2023-10-09

Expiry Date: 2028-11-06

Page: 1 of 5



EU Type Examination Certificate

No. CE 698940

Product Specification

Model:	Ambulance high risk - Powder Free Latex Gloves.
Classification:	Protective gloves for use against dangerous chemical and micro-organism hazards.
Description:	A five fingered, ambidextrous, single use powder free, non-sterile, single/double chlorinated Latex glove with textured surface and beaded cuff. Gloves are coloured blue.
PPE Category:	Category III
Size range:	XS, S, M, L, XL, XXL
Applicable Standards:	<p>The following standards:</p> <p>EN ISO 21420:2020 Protective gloves. General requirements and test methods. Incorporating Amendment 1 from ISO 21420:2020/Amd.1:2022.</p> <p>EN ISO 374-1:2016+A1:2018 Protective gloves against dangerous chemicals and micro-organisms. Terminology and performance requirements for chemical risks.</p> <p>EN ISO 374-2:2019 Protective gloves against dangerous chemicals and micro-organisms. Determination of resistance to penetration.</p> <p>EN ISO 374-4:2019 Protective gloves against dangerous chemicals and micro-organisms. Determination of resistance to degradation by chemicals</p> <p>EN ISO 374-5:2016 Protective gloves against dangerous chemicals and micro-organisms. Terminology and performance requirements for micro-organisms risks</p> <p>EN 16523-1:2015+A1:2018 Determination of material resistance to permeation by chemicals. Permeation by potentially hazardous liquid chemicals under conditions of continuous contact.</p> <p>ISO 16604:2004 Clothing for protection against contact with blood and body fluids. Determination of resistance of protective clothing materials to penetration by blood-borne pathogens. Test method using Phi-X174 Bacteriophage.</p>

First Issued: 2018-11-06

Latest Issue: 2023-10-09

Effective Date: 2023-10-09

Expiry Date: 2028-11-06

Page: 2 of 5

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated [online](#).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.

EU Type Examination Certificate

No. CE 698940

Product Specification (continued)

Performance

General requirements for gloves to EN ISO 21420:2020 Incorporating Amendment 1 from ISO 21420:2020/Amd.1:2022

Characteristic	Level
Dexterity	5
Size	XS, S, M, L, XL, XXL
pH	Pass
PAH	Pass

Terminology and performance requirements for micro-organism risks EN ISO 374-5:2016

Characteristic	Level
Protection against bacteria and fungi (Test method EN ISO 374-2:2019)	Pass
Protection against viruses (Test Method ISO 16604:2004)	Pass

First Issued: 2018-11-06

Latest Issue: 2023-10-09

Effective Date: 2023-10-09

Expiry Date: 2028-11-06

Page: 3 of 5

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated [online](#).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.

EU Type Examination Certificate

No. CE 698940

Product Specification (continued)

Resistance to chemical permeation to EN ISO 374-1:2016+A1:2018 Type B Chemical Protection (Test method EN 16523-1:2015+A1:2018)

Chemical	Permeation Level
40% Sodium Hydroxide (K)	6
30% Hydrogen Peroxide (P)	5
37% Formaldehyde (T)	6

Resistance to degradation to chemical protection EN ISO 374-4:2019

Chemical	Mean Degradation %
40% Sodium Hydroxide (K)	-40.0
30% Hydrogen Peroxide (P)	22.8
37% Formaldehyde (T)	-97.8

First Issued: 2018-11-06

Latest Issue: 2023-10-09

Effective Date: 2023-10-09

Expiry Date: 2028-11-06

Page: 4 of 5

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated [online](#).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.

EU Type Examination Certificate

No. CE 698940

Certificate Administration Details

Product initially approved on BSI certificate: CE 683312.

Certificate Amendment Record:

Issue Date	Comments	Internal BSI Project Number
November 2018	First issue	0086:18:9642991
October 2023	Amendment to update the standards to the current revisions, remove the chemicals and Certificate renewal.	2797:23:3969815

Note: The Certificate holder is responsible for ensuring that the Notified Body is advised of changes to any aspect of the overall processes utilised in the manufacture of the product, failure to do so could invalidate the Certificate in respect of product manufactured following the introduction of such changes.

Monitoring of manufactured PPE:

The validity of the Certificate for the products is also dependent on the maintenance of the EU Conformity to Type Based on Quality Assurance of the Production Process, Annex VIII (Module D), as referenced on BSI issued Certificate CE 688313.

First Issued: 2018-11-06

Latest Issue: 2023-10-09

Effective Date: 2023-10-09

Expiry Date: 2028-11-06

Page: 5 of 5

This certificate has been issued by and remains the property of BSI Group The Netherlands B.V., John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands and should be returned immediately upon request.
To check its validity telephone +31 20 3460780. An electronic certificate can be authenticated [online](#).

BSI Group The Netherlands B.V., registered in the Netherlands under number 33264284, at John M. Keynesplein 9, 1066 EP Amsterdam, The Netherlands
A member of BSI Group of Companies.